

SurPure™ Prime

Special Features:

- The 3-layer design; two layers of latex with an extra polymer coating. Features wet donning ability and enhanced durability.
- Better grip performance than the standard model.
- Very low protein content, less than 50µg/dm².
- Made from selected natural latex rubber. Easy to don and comfortable to wear.
- Penetration of Phi-X174 Bacteriophage tested with ASTM F 1671.
- Pass Regulation EU 2016-425 PPE Module B.



Product Specification Ref. 1610

Product Description	<i>Wet Donning Latex Surgical Gloves, powder-free, anatomic shape, beaded cuff, textured surface in the grip area, size embossed, sterile.</i>								
Size		5.5	6.0	6.5	7.0	7.5	8.0	8.5	9.0
Palm Width	mm	70±6	76±6	83±6	89±6	95±6	102±6	108±6	114±6
Overall Length	mm	290 ± 5							
	inch	11.42 ± 0.2							
Single Wall Thickness		Finger		Palm		Cuff			
	mm	0.22 ± 0.03		0.22 ± 0.03		0.21 ± 0.03			
	mil	8.66 ± 1.18		8.66 ± 1.18		8.26 ± 1.18			
Physical Property	Type	Before Aging			After Aging				
		Standard	MocareHealth	Standard	MocareHealth				
	Force at break (N)	9	18	9	17				
	Tensile strength (Mpa)	24 (min)	26	18 (min)	23				
	Ultimate elongation (%)	750	820	560	780				
Stress at 500% elongation (Mpa)	5.5 (max)	3.5	-	-					
Sterilization	Gamma/E-beam irradiation								
Quality Standard	<ul style="list-style-type: none"> • Conforms to ASTM F 1671, ASTM D 3577, ASTM D 5712, EN 455, EU 2016-425 PPE Module B, ISO 10282 and AS/NZS 4179. • Pinhole 1.5 / 1.0 / 0.65 optional. • CE, FDA 510K approved. 								
Packaging	50 prs / box, 6 boxes / carton								

Mocare Health Co., Ltd.

Tel : 886-4-8762747 Website : <http://www.mocarehealth.com>

Fax : 886-4-8762749 E-mail : service@mocarehealth.com

Office : No.216, Laonong Rd., Tianzhong Township, Changhua County, 52046, Taiwan (R.O.C.)



